


## Introduction

The MI007 is a passive high impedance probe, designed for use with oscilloscopes. It is calibrated to work effectively with instruments that have an input impedance of 1M Ohm, shunted by 20 pF. You can also use it with oscilloscopes that have an input capacitance of 10-50pF, if you make an adjustment to compensate. In the head of the probe is a two position slide switch which allows you to select the attenuation you require: either X1 or X10.

## Frequency compensation adjustment

Connect the probe to an oscilloscope and a 1KHz square waveform source. Move the attenuation switch on the probe to the X10 position. Set the oscilloscope to display two to three cycles, and two to six vertical divisions. Carefully adjust the compensation trimmer tool on the head of the probe until the waves displayed on the oscilloscope are perfectly square.



**Pico Technology Ltd**

Tel +44 (0)1480 396395 Fax +44 (0)1480 396395  
www.picotech.com E-mail: sales@picotech.com


## Specifications:

### Position X10

Attenuation Ratio	10:1
Bandwidth	DC to 60MHz
Rise time	5.8ns
Input Resistance	10M when used with oscilloscopes with 1M input
Input Capacitance	Approx. 15pF
Compensation Range	10 to 50pF
Working Voltage	600VDC incl. Peak AC derating with frequency (see Fig.1)

### Position X1

Attenuation Ratio	1:1
Bandwidth	DC to 15MHz
Rise Time	23.3ns
Input Resistance	1M(Oscilloscope Input)
Input Capacitance	46pF plus oscilloscope capacitance
Working Voltage	600V DC incl. Peak AC Derating with frequency
Safety	conformed IEC-1010 CAT II
Cable Length	1.2Meter



Update:22-04-2011

## **X-ON Electronics**

Largest Supplier of Electrical and Electronic Components

*Click to view similar products for [pico manufacturer](#):*

Other Similar products are found below :

[MI136](#) [PA149](#) [PICO CONNECT 442](#) [SE050](#) [SE055](#) [SE056](#) [SE057](#) [SE058](#) [PICOCONNECT 910 KIT](#) [TA328](#) [TA155](#) [MI078](#) [81170](#) [SE049](#)  
[SE054](#) [MI106](#) [TA384](#) [TA385](#) [100501](#) [SE053](#) [44740M2735659](#) [PICOLOG 1216](#) [TA139](#) [TA313](#) [TC-08 BUNDLE](#) [PICOSCOPE 5442D](#)  
[TA050](#) [TA018](#) [42345M2735662](#) [TA089](#) [75005](#) [TA090](#) [PS010](#) [TA058](#) [SE028](#) [TC-08](#) [PICOSCOPE 2204A](#)