



Macromolecule Humidity Sensor

(Model: MS-Z2)

Manual

Zhengzhou Winsen Electronics Technology Co., Ltd

Statement

This manual copyright belongs to Zhengzhou Winsen Electronics Technology Co., LTD. Without the written permission, any part of this manual shall not be copied, translated, stored in database or retrieval system, also can't spread through electronic, copying, record ways.

Thanks for purchasing our product. In order to let customers use it better and reduce the faults caused by misuse, please read the manual carefully and operate it correctly in accordance with the instructions. If users disobey the terms or remove, disassemble, change the components inside of the sensor, we shall not be responsible for the loss.

The specific such as color, appearance, sizes &etc, please in kind prevail.

We are devoting ourselves to products development and technical innovation, so we reserve the right to improve the products without notice. Please confirm it is the valid version before using this manual. At the same time, users' comments on optimized using way are welcome.

Please keep the manual properly, in order to get help if you have questions during the usage in the future.

Zhengzhou Winsen Electronics Technology CO., LTD

MS-Z2 Macromolecule Humidity Sensor

Overview

In wet conditions, water molecules are adsorbed by polar group on the surface of material. And as the humidity increases, the quantity of water molecules will be changed accordingly. The adsorbed water is gradually condensing and coming into being liquid, which is electrolyte solution with current channel quality.

With the humidity increasing, macromolecule will swell, interior free volume will be bigger, carrier will be increased and the activated energy of macromolecule polyelectrolyte counter-ions will decrease, drift mobility will increase and impedance will decrease. And then when humidity decreases, water molecules are released from ion polymer and the resistor of material will increase. The environment humidity can be monitored through testing the impedance.

Features


- Wide humidity detected range
- Fast response
- Small Humidity hysteresis error
- Simple manufacture
- Easy integration

Application


Humidity sensor, as an important chemical sensor, which is widely used in fields of warehousing, industry production, and process control, environmental monitoring, home appliances and meteorology etc.

Technical specification Basic testing circuit


	Temperature	Humidity
Working range	0~60°C	10~90%RH
Storage range	-25~70°C	≤95%RH(non condensation)
Detection range		10~90%RH(0~60°C)
Rated voltage	1.5V AC(MAX, sine wave)	
Rated power	0.2mW(MAX, Rated voltage)	
Working frequency	500Hz~2kHz	
Nominal value &range	31 (21~50) KΩ(60%RH, 25°C) 23(15~35)KΩ(60%RH, 25°C)	
Temperature Character	≤0.5%RH/°C	
Hysteresis	±2%RH	
Response time	Moisture absorption: ≤20s Dehumidification: ≤40s	
Stability	2%RH/year	
Accuracy	3%RH	




Sensitivity & Humidity Hysteresis Error Curve



Temperature & Humidity Characteristic



Application Circuit



Note

- To avoid Polarization, the voltage or current driving the sensor should be AC
- Please using LCR AC bridge to measure, and multi-meter is prohibited.
- Avoid Water coagulation
- Putting the sensor under the High causticity place is prohibited
- Recommend storage conditions:
 Temperature: 10°C~ 40°C
 Humidity: under 60%RH

Zhengzhou Winsen Electronics Technology Co., Ltd

Add: No.299, Jinsuo Road, National Hi-Tech Zone,
Zhengzhou 450001 China

Tel: +86-371-67169097/67169670

Fax: +86-371-60932988

E-mail: sales@winsensor.com

Website: www.winsen-sensor.com

X-ON Electronics

Largest Supplier of Electrical and Electronic Components

Click to view similar products for [Environmental Test Equipment](#) category:

Click to view products by [WINSEN](#) manufacturer:

Other Similar products are found below :

[CW40](#) [F150C10E3DRT](#) [F150LTC20](#) [F150CD10E2](#) [F150L75](#) [F150LRS](#) [S-11](#) [FLUKE-LDG](#) [382153](#) [FM100](#) [RH210](#) [382253](#) [MO290-P](#)
[GEO-CABLE-REEL-50M](#) [T197914](#) [RD300-L](#) [IR11BD](#) [IR11GM](#) [IR12GM](#) [IR21BD](#) [IR31CE](#) [IR32BC](#) [IR33BC](#) [MP7217TC](#) [NGM-1](#) [SGX-](#)
[7NH3](#) [UT381](#) [F150-SLC50](#) [AW-CO-1000](#) [AW-NmHc-100](#) [3.000.401](#) [AX-7535](#) [CS-9S6SS-A](#) [P 2800 A](#) [P 2801](#) [P 2802](#) [P 5039](#) [P 5130](#) [P](#)
[5055](#) [P 5060](#) [P 5065](#) [P 5086](#) [P 5090](#) [P 5110](#) [P 5115](#) [P 5135](#) [P 5140](#) [P 5145](#) [P 5150](#) [P 5160](#)