

## Operation Sheet

**EXTECH**<sup>®</sup>  
INSTRUMENTS  
A FLIR COMPANY

# Heavy Duty Continuity Tester

## Model ET40



### Features:



Ideal for checking continuity of non-energized components, fuses, switches, relays, wiring and circuit boards

Features 30" (76cm) Test lead with alligator clip

### Continuity Testing

**WARNING:** To avoid electric shock, never measure continuity on circuits or wires that have voltage on them.

1. Install one AAA battery.
2. Remove all power from the circuit to be tested.
3. Remove the plastic probe tip.
4. Touch the probe tip to one side of the device and the alligator clip to the other side
5. The bulb will light if there is continuity.



**CAUTION:** Use extreme caution when checking electrical circuits to avoid injury due to electrical shocks. Extech Instruments assumes basic knowledge of electricity on the part of the user and is not responsible for any injury or damage due to improper use of this tester.

Copyright © 2012 Extech Instruments Corporation (a FLIR company)  
All rights reserved including the right of reproduction in whole or in part in any form.  
[www.extech.com](http://www.extech.com)

## **X-ON Electronics**

Largest Supplier of Electrical and Electronic Components

*Click to view similar products for [extech](#) manufacturer:*

Other Similar products are found below :

[PT400](#) [EX365](#) [MO260](#) [MO280](#) [PH113](#) [HD700](#) [HD755](#) [140000](#) [AN300](#) [149116](#) [149118](#) [DV40](#) [EA11A](#) [EMF450](#) [BT100](#) [SDL600](#) [EX363](#)  
[SW810A](#) [AN200](#) [AN320](#) [PH110](#) [PH220](#) [PRC15](#) [PRC20](#) [PT-30](#) [HD500](#) [HDV540](#) [RH490](#) [RH522](#) [380405](#) [380941](#) [380975](#) [382270](#) [MA145](#)  
[MA1500](#) [MA3010](#) [407736](#) [412300](#) [412355A](#) [42275](#) [42540](#) [TL726](#) [TL807C](#) [45118](#) [461895](#) [380260](#) [380363](#) [380462](#) [380926](#) [381676A](#)