

## METRAHIT 2+ Universal TRMS Multimeter

3-349-476-03 11/6.19

- Resolution: 100 μV, 100 mΩ, 10 μA
- TRMS measurement
- Precision temperature measurement
- Automatic and manual measuring range selection
- Backlit digital display with additional analog scaleMeasured value memory,
- HOLD, MIN / MAX valueOverload and blown fuse indicators
- IP 40 protection
- Protective rubber cover
- 3 year guarantee
- DAkkS calibration certificate included as a standard feature









### Features

#### Automatic Blocking Sockets (ABS) \*

Automatic blocking sockets prevent incorrect connection of measurement cables and inadvertent selection of the wrong measured quantity. This significantly reduces danger to the user, the instrument and the system under test, and eliminates it entirely in many cases.

#### Automatic / Manual Measuring Range Selection

Measured quantities are selected with the rotary switch. The measuring range is automatically matched to measured values. The measuring range can be selected manually as well with the help of the AUTO/MAN key.

#### Display of Negative Values at the Analog Scale

Negative values are also displayed at the analog scale for zerofrequency quantities, allowing for observation of measured quantity fluctuation around the zero-point.

#### Storage of Measured Values

By pressing the **HOLD**/MIN/MAX key, the currently displayed measurement value can be "frozen" in the display. The minimum and maximum values which were present at the input of the measuring instrument after activation of the MIN/MAX mode can be selectively "retained" with the MIN/ MAX function. The most important application is the determination of the minimum or maximum value during long-term observation of measurement quantities. MIN/MAX has no effect on the analog display; it continues to display the current measurement value.

#### **Continuity Test**

Allows for the detection of short-circuits and interrupted conductors. In addition to displaying test results, an acoustic signal can also be generated if desired.

#### **Power Saving Circuit**

The device is switched off automatically if the measured value remains unchanged for a period of approximately 10 minutes, and if none of the controls are activated during this time. Automatic shutdown can be deactivated.

#### **Protective Cover for Harsh Conditions**

The instrument is protected against damage in the event of impacts or dropping by means of a soft rubber cover with tilt stand. The rubber material also assures that the instrument does not wander if it is set up on a vibrating surface.

#### **RMS Measurement with Distorted Waveshapes**

The measuring method applied allows for RMS measurement for alternating signals (AC) in voltage and current measurement, independent of the waveshape up to 1 kHz (for non-sinusoidal signals as well).

\* Patented (patent no. EP 1801 598 and US 7,439,725)

## METRAHIT 2+ **Universal Multimeter**

## **Applicable Regulations and Standards**

IEC 61010-1/EN 61010-1/ VDE 0411-1	Safety requirements for electrical equipment for measurement, control and laboratory use
EN 60529 VDE 0470, Part 1	Test instruments and test procedures Protection provided by enclosures (IP code)
IEC 61326/EN 61326	Electromagnetic compatibility (EMC)

### Voluntary Manufacturer's Guarantee

36 months for material and workmanship 1 ... 3 years for calibration (depending on application)

## **Characteristic Values**

Meas.	Magguring Dange	Reso-	Input Im	pedance	under Referer	Max. Resolution ace Conditions	Overload	Capacity	Meas.										
Function	Measuring Range	lution			±( % rdg. + d)	±( % rdg. + d)			Function										
		6000		$\sim$		~5)	Value	Time											
	600 mV	100 µV	$10 \text{ M}\Omega$ // < $40 \text{ pF}$	8.1 MΩ // 50 pF	0.5 + 5		600 V												
v	6 V	1 mV	$5.2 \text{ M}\Omega$ // < 40 pF	4.6 MΩ // 50 pF	0.5 + 5	1+5	DC AC	Cont.	v										
v	60 V	10 mV	$5 \text{ M}\Omega$ // < 40 pF	4.4 MΩ // 50 pF	0.5 + 5	1+5	eff	COIIL.	v										
	600 V	100 mV	$5 \text{ M}\Omega // < 40 \text{ pF}$	4.4 MΩ // 50 pF	0.5 + 5		Sinus												
			Voltage drop at a	oprox. range limit															
				$\sim$		~5)													
	60 mA	10 µA	100 mV	100 mV	1.0 + 5 (> 10 D)	1.5 + 5 (> 10 D)	1.0 A	Cont.											
Α	600 mA	100 µA	700 mV	700 mV	1.0 + 5	1.5 + 5 (> 10 D)	1.0 A	COIIL.	A										
A	6 A	1 mA	200 mV	200 mV	1.0 + 5 (> 10 D)	1.5 + 5 (> 10 D)	10 A <sup>4)</sup>	Cont.	~										
	10 A	10 mA	300 mV	300 mV	1.0 + 5	1.5 + 5 (> 10 D)	IUA	COIIL.											
			Open-circuit voltage	Meas. current at range limit	±( % rc	g. + d)													
	600 Ω	$100\text{m}\Omega$	max. 1 V	max. 250 μA	1 + 5 <sup>2)</sup>														
	6 kΩ	1 Ω	max. 1 V	max. 100 μA	0.7 + 3		0001/												
Ω	60 kΩ	10 Ω	max. 1 V	max. 12 μA	0.7 + 3		600 V DC			Ω									
22	600 kΩ	100 Ω	max. 1 V	max. 1.2 μA	0.7 + 3		AC max. 10 s eff Sinus		52										
	6 MΩ	1 kΩ	max. 1 V	max. 120 nA	0.7 + 3														
	40 MΩ	10 kΩ	max. 1 V	max. 50 nA	2.0 + 3		Sinus												
₩	2 V	1 mV	max. 3 V		1 + 5				₩										
					±( % ro	g. + d)													
<b>L</b> ))	600 Ω	0.1 Ω	max. 1 V		1 + 5		600 V DC AC	max. 10 s	<b>u</b> ())										
					±( % rd	g. + K)													
°C	ТҮР К -50.0 + 400.0 °С	0.1 °C			1.0 + 5	K <sup>3)</sup>	600 V DC/AC eff Sinus	max. 10 s	°C										
					±( % rc	g. + d)													
Hz	100 Hz	0.1 Hz			0.1 + 2		600 V <sup>6)</sup>		Hz										
пг	1000 Hz	1 Hz			0.1 + 2		000 V "/		П2										

<sup>1)</sup> At 0 to + 40 °C

2) With zero balancing, or + 35 digits without zero balancing

With 2010 Data to 1.9, 1
Without sensor
12 A for 5 min, 16 A for 30 s
12 A for 5 min, 16 A for 30 s

<sup>47</sup> 12 A for 5 min, 16 A for 30 s
<sup>50</sup> 1 ... 35 d from the zero point due to TRMS converter when probe tips are short-circuited
6) power limiting: frequency x voltage max. 3 x 10<sup>6</sup> V x Hz for U > 100 V

#### Key

rdg. = reading (measured value) d = digit

#### **Reference Conditions**

Ambient temperature	+ 23 °C ± 2 K
Relative humidity	40 60%
Measured quantity frequency Measured quantity	45 65 Hz
waveshape	Sinusoidal
Battery voltage	$3~V~\pm 0.1~V$

# METRAHIT 2+ Universal Multimeter

#### Influencing Quantities and Influence Error

Influencing Quantity	Sphere of Influence	Measured Quantity / Measuring Range	Influence Error <sup>1)</sup> ±( % rdg. + digits)
		600 mV <del></del>	1.0 + 3
		6 600 V <del></del>	0.15 + 1
		٧~	0.4 + 2
		60 mA 600 mA <del></del>	0.5 + 1
	0 °C +21 °C	6 A/10 A <del></del>	0.5 + 1
Temperature	and	$A \sim$	0.75 + 1
	+25 °C +40 °C	0 Ω <sup>2)</sup>	0.15 + 2
		600 Ω	0.25 + 2
		6 kΩ 6 MΩ	0.15 + 1
		40 MΩ	1.0 + 1
		− 50 + 200 °C	1 K + 2
		+ 200 + 400 °C	1 + 2
	> 30 Hz 45 Hz	$A \sim$	2.0 + 10
	> 65 Hz 1 kHz	60 / 600 mA / 6 A	1.5 + 10
Measured	> 03 HZ T KHZ	10 A	2 + 10
Quantity	> 20 Hz /5 Hz	600 mV	3 + 10
Frequency	> 30 Hz 45 Hz	6 / 60 /600 V	2.5 + 10
	> 65 Hz 500 Hz	600 mV	3.5 + 20
	> 65 Hz 800 Hz	6 / 60 V	2.5 + 10

Influen- cing Quantity	Sphere of Influence	Measured Quantity / Measuring Range	Influence Error
		V <del></del>	± 2 Digits
		$V \sim$	$\pm$ 4 Digits
Battery	<b>⊣⊢</b> <sup>3)</sup> < 2.9 V	A	± 4 Digits
Voltage >	> 3.1 V 3.6 V	$A \sim$	± 6 Digits
		60 Ω / 600 Ω / °C	± 4 Digits
		6 kΩ 40 MΩ	± 3 Digits
Relative Humidity	75% 3 days Instrument off	V <u>~</u> A <u>~</u> Ω °C	1 x intrinsic error
HOLD	_		± 1 Digits
MIN / MAX	—	V $\simeq$ , A $\simeq$	± 2 Digits

<sup>1)</sup> For temperature: specified error valid starting with temperature changes as of 10 K. For frequency: specified error valid starting with display values as of 300 digits.

2) With zero balancing
 3) After the + symbol appears at the display

Influencing Quantity	Sphere of Influence	Measuring Ranges	Damping
	Interference quantity max. 600 V $\sim$	V <del></del>	> 120 dB
Common Mode Interference Voltage	Interference quantity max. 600 V $\sim$	6 V $\sim$ , 60 V $\sim$	> 80 dB
interference vertage	50 Hz, 60 Hz sine	600 V $\sim$	> 70 dB
Series Mode Interference Voltage	Interference quantity: V ~, respective nominal value of the measuring range, max. 600 V ~, 50 Hz, 60 Hz sine	V	> 50 dB
	Interference quantity max. 600 V —	V~	> 110 dB

#### Crestfaktor CF

Test signal: Rectangle 55 Hz, no DC component



Influencing Quantity	Sphere of Influence	Measured Quantity / Measuring Range	Influence Error
Crest factor CF	$1.5 < CF \le 2$	6 V. 60 V. 600 V ~	±1 % rdg.
GIEST INCLUI OF	$2 < CF \le 4$	0 v, 00 v, 000 v $\sim$	±5 % rdg.

The admissible crest factor CF of the alternating quantity to be measured depends on the display value.

Crest factor 4 at the end of range, it is increased accordingly when the range is reduced. However, due to input protection, voltage is limited to 1000 V, therefore the admissible crest factor in the 600 V ranges is half as high.

Power limiting: voltage x frequency max.  $3 \times 10^{6}$  V x Hz.

#### Response Time (after manual range selection)

Measured Quantity /	Response Time		Measured Quantity
Measuring Range	Analog Display	Digital Display	Step Function
V <del>,</del> V ~, A <del>,</del> A ~	0.7 s	1.5 s	from 0 to 80% of the upper range limit
600 Ω 6 MΩ	1.5 s	2 s	
40 MΩ	4 s	5 s	from ∞ to 50% of the upper range limit
*	_	1.5 s	
°C	_	max. 1 3 s	from 0 to 50% of the upper range limit

#### Display

LCD panel (65 mm x 30 mm) with analog and digital display including unit of measure, type of current and various special functions. Background illumination the **ON** / **OFF** key, and is switched off automatically after approximately 1.

#### minute. <u>Analog</u>:

<u>Analog</u> :	
Display	LCD scale with pointer
Scale length	55 mm in all ranges
Scaling	$0 \hdots \pm 60$ with 61 scale divisions in all
	ranges
Polarity display	With automatic switching
Overflow display	Triangle
Measuring rate	20 measurements per second
<u>Digital:</u>	
Display / char. height	7-segment characters / 15 mm
Number of places	$3^{6}/_{7}$ -place $\leq$ , 6000 steps
Overflow display	"0.L" appears
Polarity display	"-" sign is displayed if plus pole is connected to $\bot$
Measuring rate	2 measurements per second

# METRAHIT 2+ Universal TRMS Multimeter

#### **Power Supply**

Battery	2 x 1.5 V AA size batteries, alkaline manganese per IEC LR6 or equivalent rechargeable NiCd battery
Service life	With alkaline manganese: approx. 750 hours for V , A approx. 200 hours for V $\sim$ , A $\sim$
Battery test	<ul> <li>Is displayed automatically if battery voltage drops to below approximately 2.1 V.</li> </ul>
Electrical Safety	

Safety class	II per IEC 61010-1:2010/EN 61010-1:2010/ VDE 0411-1:2011
Measuring category	CAT III
Nominal voltage	600 V
Pollution degree	2
Test voltage	5.2 kV~ per IEC 61010-1/EN 61010-1

#### **Electromagnetic Compatibility (EMC)**

Interference emission	EN 61326-1: 2013 class B
Interference immunity	EN 61326-1: 2013
	EN 61326-1-2: 2013

#### Fuses

Fuse links for all ranges

up to 600 mA	FF 1.6 A/1000 V, 6.3 mm x 32 mm, switching capacity: 10 kA at 1000 V~ with ohmic load, protects all current measuring ranges up to 600 mA in combination with power diodes
Fuse links for all	
ranges up to 10 A	FF 10 A/1000 V, 10 mm x 38 mm, switching capacity: 30 kA at 1000 V with ohmic load, protects 6 A and 10 A ranges to 1000 V

#### Ambient Conditions

Accuracy range	0 °C + 40 °C
Operating temp.	−10 °C + 50 °C
Storage temperature	- 25 °C + 70 °C without batteries
Relative humidity	45 75%, no condensation allowed
Elevation	to 2000 m

#### **Mechanical Design**

Protection	IP 40, IP 20 at the connector jacks per DIN VDE 0470, part 1 / EN 60529
Dimensions	84 mm x 195 mm x 35 mm
Weight	Approx. 350 gr. with battery

#### Prepared in Germany • Subject to change without notice • A pdf version is available on the Internet



GMC-I Messtechnik GmbH Südwestpark 15 90449 Nürnberg • Germany Phone +49 911 8602-111 Fax +49 911 8602-777 E-Mail info@gossenmetrawatt.com www.gossenmetrawatt.com

### **Standard Equipment**

- 1 TRMS-Digital-Multimeter
- 1 protective rubber holster
- 2 2 x 1.5 V AA size batteries
- 1 set of measurement cables KS17-ONE
- 1 DAkkS calibration certificate
- 1 Short-form operating instructions

Detailed operating instructions are available on our website www.gossenmetrawatt.com.

### **Order Information**

Description	Туре	Article Number
Analog-digital multimeter standard equipment see above	METRAHIT 2+	M205A
Accessories		
Fast reacting surface temperature sensor, type K (NiCr-Ni) –50 +400 °C	TF400SURFACE	Z102E
Clip-on current transformer, 30 mA 150 A~, 1000:1, ±2.5 %, 1 mA/A	WZ12D	Z219D
Carrying pouch	F829	GTZ3301000R0003
Imitation leather carrying pouch for one METRAHit <sup>®</sup> and accessories	F836	GTZ3302000R0001
Hard case for 1 METRAHit <sup>®</sup> and accessories	HC20	Z113A
Hard case for two METRAHit <sup>®</sup> , adapter and accessories	HC30	Z113A
Fuses (pack of 10)	FF 1.6 A / 1000 V	Z109C
Fuses (pack of 10)	FF 10 A / 1000 V	Z109L

For additional information on accessories, please refer to

- our "Measuring Instruments and Testers" catalogue
- our website www.gossenmetrawatt.com

## **X-ON Electronics**

Largest Supplier of Electrical and Electronic Components

Click to view similar products for Digital Multimeters category:

Click to view products by Gossen Metrawatt manufacturer:

Other Similar products are found below :

 6111-517
 FS881
 40705X
 C.A 6133
 LAUNCH KIT
 P 1020
 A
 SEFRAM7303
 BS K-CLIP
 19290
 DM285-FLEX-KIT
 IBT6K
 1000-219

 1001-613
 1006-969
 1008-221
 1012-597
 1013-099
 30XR
 34XR
 35XP
 TESTO 745
 0590
 760-2
 0590
 7602
 TESTO 760-3

 0590
 7603
 440012
 AX-155
 AX-174
 AX-178
 AX-18B
 AX-190A
 AX-503
 AX-507B
 AX-594
 AX-LCR42A
 AX-MS811
 AX-MS8250
 AX 

 PDM01
 AX-T520
 AX-T901
 AX-T903
 BAT-250-EUR
 BM5255
 BM8055
 BM8075
 BM8175
 BM8275
 BM8295
 BM8595

 BM867S
 BM907S
 33XR
 33XR
 BM5255
 BM8055
 BM8075
 BM8275
 BM8295
 BM8595