

# LOOP/PFC/PSC TESTER KEW 4140



#### Specifications

Lo	op Impedance					
	Function	L-PE ATT OFF  230V (50/60Hz)  100~280V (45~65Hz)		L-N/L-L		
	Rated voltage			L-N: 230V (50/60Hz) L-L: 400V (50/60Hz)		
	Operating Voltage				100~500V (45~65Hz) 20Ω 20Ω: 6A/20ms  L-N: ±3%rdg±4dgt L-L: ±3%rdg±8dgt	
	Range(Auto-Ranging)	20/200/2000Ω 20/200/2000Ω (L-N<20Ω)		-N<20Ω)		
	Nominal Test Current at 0Ω External Loop: Magnitude/Duration at 230V	20Ω: 6A/40ms 200Ω: 2A/20ms 2000Ω: 15mA/500ms	00Ω: 2A/20ms N-PE: 10mA/approx.5s 000Ω: 15mA/500ms			
	Accuracy	±3%rdg±4dgt (*1)				
PF	C(L-PE)/PSC(L-N/L-L) (*2)					
	Function	PFC (ATT)			PSC	
	Rated voltage	230V (50/60Hz)		L-N: 230V (50/60Hz) L-L: 400V (50/60Hz)		
	Operating Voltage	100~280V (45~65Hz)		100~500V (45~65Hz)		
	Range(Auto-Ranging)	2000A/20kA	2000A/20kA (L-N<20Ω)		2000A/20kA	
	Nominal Test Current at 0Ω External Loop: Magnitude/Duration at 230V	20Ω: 6A/40ms 200Ω: 2A/20ms 2000Ω: 15mA/500ms	L-N: 6A/60ms N-PE: 10mA/approx.5s		20Ω: 6A/20ms	
Ph	ase Rotation					
	Operating Voltage	50~500V, 45~65Hz	0~500V, 45~65Hz			
	Remarks		Correct phase sequence : displayed "1.2.3" and ♥ mark Reversed phase sequence : displayed "3.2.1" and ♥ mark			
/ol	ts					
	Function Volts		Frequency			
	Measuring range	0~500V		45~65Hz		
	Accuracy	±2%rdg±4dgt		±0.5%rdg±2dgt		
Applied standards		IEC 61010-1 CATII 300V (500V L to L) IEC 61010-031, IEC 61557-1,3,7,10, IEC 60529 (IP54), IEC 61326 (EMC)				
0	wer source	1.5V AA batteries × 6 *Use of alkaline batteries (LR6) is recommended.				
Dir	nensions	84(L) × 184(W) × 133(D)mm				
Νe	eight	860g (including batteries.)				
<del>ا</del> د	cessories included	7218 (Main cord) (13), 7246 (Distribution board test lead), 9155 (Shoulder strap), 9156 (Soft case) LR6 (Battery) × 6, Instruction manual, Calibration certificate				

<sup>\*1:</sup> Accuracy of L-N LOOP displayed on the Sub Display is synchronized with the one at L-N/L-L function. 
\*2: PSC/PFC Accuracy is derived from measured loop impedance specification and measured voltage specification. 
\*3: 7187 (Power cord [UK]), 7221 (Power cord [AU])

#### **EVERYTHING YOU NEED ....**

KEW 4140 comes complete with everything you need for Loop, PFC, PSC and Phase rotation testing on an electrical installation according to the IEC 60364-6 (or its national equivalent standard). A full set of new safety test leads is included, they have an overmould for maximum grip and are extremely robust. The instrument also comes in a soft carry case, a quick reference guide locates on the instrument body, a strap belt and it is supplied with a calibration certificate.





Please read the "Safety Warnings" in the instruction manual supplied with the instrument thoroughly and completely Safety Warnings: If the Institution and Supplied with the Institution and to operate the instrument on a correct power supply and voltage rating marked on each instrument.

#### For inquires or orders:



### **KYORITSU ELECTRICAL INSTRUMENTS** WORKS, LTD.

No.5-20, Nakane 2-chome, Meguro-ku, Tokyo, 152-0031 Japan Phone:81-3-3723-0131 Fax:81-3-3723-0152

E-mail:info@kew-ltd.co.jp Factory: Ehime, Japan

http://www.kew-ltd.co.jp





## **X-ON Electronics**

Largest Supplier of Electrical and Electronic Components

Click to view similar products for Environmental Test Equipment category:

Click to view products by Kyoritsu manufacturer:

Other Similar products are found below:

CW40 F150C10E3DRT F150LTC20 F150CD10E2 F150L75 F150LRS 2837806 4328074 4366444 S-11 382153 700PTP-1 GEO-CABLE-REEL-50M H115 H300 F150-SLC50 TM-192 TM-192D TM-414 EL-P-TC-T-SURFACE UT301C+ HS115 12227021 13123602 13123661 12255930 12983030 30171091 30198712 30198755 30209323 12228657 12229067 12230090 7358958 12231398 12231487 7390674 7390682 12231550 7391840 7391939 7391955 7391972 7391998 7392056 7392099 7395489 7400058 7400394