



All dimensions are in mm; tolerances according to ISO 2768 m-H

Interface

According to

IEC 61169-4, EN 122190, DIN 47223

Contents and Documentation

This kit is delivered with

- **Standard Definitions Card**
Printed Standard Definitions that can be used on nearly all Vector Network Analyzers
- **Test Results Documentation**
- **Lanyard**
- **Hard Shell Case**
- **Protection Caps**

Material and plating

Connector parts

Center conductor
Outer conductor
Body
Dielectric
Substrate

Material

CuBe
Brass
Brass
PP
Al₂O₃

Plating

Gold, min. 1.27 µm, over nickel
Flash white bronze over silver(e.g. Optargen®)
powder coated

Electrical data

Frequency range DC to 4 GHz

Open

Return loss ≤ 0.15 dB, DC to 4 GHz
 Error from nominal phase¹ $\leq 3.0^\circ$, DC to 4 GHz

Short

Return loss ≤ 0.15 dB, DC to 4 GHz
 Error from nominal phase² $\leq 3.0^\circ$, DC to 4 GHz

Load

Return loss ≥ 40 dB, DC to 2.5 GHz
 ≥ 38 dB, 2.5 GHz to 4 GHz
 DC-Resistance $50 \Omega \pm 0.5 \Omega$
 Power handling (at 25 °C, sea level) ≤ 1.0 W, derate by 0.01 W/K

¹ The nominal phase is defined by the Offset Delay, the Offset Loss and the Fringing Capacitances

² The nominal phase is defined by the Offset Delay, the Offset Loss and the Short Inductance

Mechanical data

Mating cycles ≥ 500
 Maximum torque 30 Nm
 Recommended torque 2.26 Nm
 Gauge 1.77 mm to 2.07 mm

General standard definitions

For proper operation the vector network analyzer (VNA) needs a model describing the electrical behaviour of this calibration standard. The different models, units, and terms used will depend on the VNA type and they will have to be entered into the VNA. All values are based on typical geometry and plating.

Open

Offset Z_o / Impedance / Z_o 50 Ω
 Offset Delay 87.394 ps
 Length (electrical) / Offset Length 26.20 mm
 Offset Loss 0.50 G Ω /s
 Loss 0.0076 dB/ $\sqrt{\text{GHz}}$
 Fringing Capacitances $C_0 = 177.000 \times 10^{-15}$ F / 177.000 fF
 $C_1 = 7200.00 \times 10^{-27}$ F/Hz / 7.20000 fF /GHz
 $C_2 = -3300.00 \times 10^{-36}$ F/Hz² / -3.30000 fF /GHz²
 $C_3 = 386.000 \times 10^{-45}$ F/Hz³ / 0.38600 fF /GHz³

Dieses Dokument ist urheberrechtlich geschützt • This document is protected by copyright • Rosenberger Hochfrequenztechnik GmbH & Co. KG

RF_35/09.14/6.2

Short

| | | | |
|-------------------------------------|--|---|----------------------------|
| Offset Z_0 / Impedance / Z_0 | 50 Ω | | |
| Offset Delay | 96.734 ps | | |
| Length (electrical) / Offset Length | 29.00 mm | | |
| Offset Loss | 0.50 G Ω /s | | |
| Loss | 0.0084 dB/ $\sqrt{\text{GHz}}$ | | |
| Short Inductance | $L_0 = 0.0000 \times 10^{-12}$ H | / | 0.0000 pH |
| | $L_1 = 0.0000 \times 10^{-24}$ H/Hz | / | 0.0000 pH/GHz |
| | $L_2 = 0.0000 \times 10^{-33}$ H/Hz ² | / | 0.0000 pH/GHz ² |
| | $L_3 = 0.0000 \times 10^{-42}$ H/Hz ³ | / | 0.0000 pH/GHz ³ |

Load

| | |
|-------------------------------------|--------------------------------|
| Offset Z_0 / Impedance / Z_0 | 50 Ω |
| Offset Delay | 0.0000 ps |
| Length (electrical) / Offset Length | 0.000 mm |
| Offset Loss | 0.00 G Ω /s |
| Loss | 0.0000 dB/ $\sqrt{\text{GHz}}$ |

Environmental data

| | |
|--|-------------------|
| Operating temperature range ³ | 0 °C to +50 °C |
| Storage temperature range | - 55 °C to +90 °C |
| RoHS | compliant |

³ Temperature range over which these specifications are valid.

Declaration of documentation

Standard delivery for this kit includes Test Results. The documentation issued reports which quantities were tested individually, traceable to national / international standards. Model based standard definitions of the calibration standards are reported in Agilent / Keysight, Rohde & Schwarz and Anritsu compatible VNA format.

Inspection interval

| | |
|----------------|-----------|
| Recommendation | 12 months |
|----------------|-----------|

Packing

| | |
|----------|--------------|
| Standard | 1 pce in bag |
| Weight | 310 g/pce |

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.

| | | | | | | | |
|----------------|----------|---------------|----------|------|---------------------------|------------------|----------|
| Draft | Date | Approved | Date | Rev. | Engineering change number | Name | Date |
| Kerstin Herzog | 10.07.06 | Markus Müller | 06.05.20 | i00 | 19-2083 | Marion Striegler | 06.05.20 |

| | | | | | | | |
|--|--|--|--|---|--|--|---------------|
| Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de | | | | Tel. : +49 8684 18-0 Email : info@rosenberger.de | | | Page 3 / 3 |
|--|--|--|--|---|--|--|---------------|

X-ON Electronics

Largest Supplier of Electrical and Electronic Components

Click to view similar products for [Calibration Equipment](#) category:

Click to view products by [Rosenberger](#) manufacturer:

Other Similar products are found below :

[700P27](#) [AX-C830](#) [AKT8R112100](#) [AKT4R1121001](#) [AKT4R1122001](#) [AR604/P](#) [FLUKE-700G08](#) [461750](#) [AX-C850](#) [RT-820](#) [FLUKE 709](#)
[FLUKE-805](#) [AKT4R111100](#) [AKT4R111200](#) [AKT7111100J](#) [AKT7212100J](#) [AKT8R111100](#) [REF2510R](#) [K39-LCRR](#) [412355A](#) [461831](#)
[461895](#) [461920](#) [480400](#) [480403](#) [PRC15](#) [PRC20](#) [PRC30](#) [700P02](#) [700P27EX](#) [700P29](#) [FLUKE-714B/EN](#) [FLUKE-789](#) [FLUKE BP7235](#)
[U086](#) [U087](#) [03K30R-MSOTS3](#) [18CK010-150](#) [19CK010-150](#) [P5514-70M](#) [P5515-140M](#) [12890](#) [TACH-10](#) [461950](#) [461825](#) [461830](#) [461891](#)
[350456](#) [CAL73](#) [DCS025](#)