

# **Datasheet SFM3019 series**

# Digital Mass Flow Meter



## **Product Summary**

SFM3019 is Sensirion's solution on providing high volumes of flow meters for medical ventilators in order to support the global pandemic situation due to the Corona virus outbreak. It relies on a modular approach of assembling approved parts with good availability to Sensirion and a high-volume through-put manufacturing.

The housing geometry is based on the SFM3000 flow meter; hence it remains mechanically compatible to the SFM3000.

A flow range from -10slm to +240slm and one directional flow supports high flow applications while still allowing to detect back flows. To minimize pressure-drop the SFM3019 is equipped with one mesh at the inlet.

### Key characteristics at a glance

- Flow range up to 240slm
- Calibrated for O<sub>2</sub>, air and mixtures thereof
- I<sup>2</sup>C interface
- Low pressure drop across the sensor
- No recalibration needed

## Benefits of Sensirion's CMOSens® Technology

- Scalability
- High reliability and long-term stability
- Best signal-to-noise ratio
- Industry-proven technology with a track record of more than 15 years
- Designed for mass production
- High process capability

#### Content

1.	Ordering Information	2
2.	Specifications	2
3.	Measurement Mode	5
4.	Digital Interface Description	7
5.	Package Outline	14
6.	Soldering	14
7.	Label information	15
8.	Shipping Package	15
9.	Revision History	15
10.	Important Notices	16



# 1. Ordering Information

Use the part names and item numbers shown in the following table when ordering the SFM3019. For the latest product information and local distributors, visit <a href="https://www.sensirion.com">www.sensirion.com</a>.

Part name	Description	Order number
SFM3019	-10 to 240slm range, with cap	3.000.400

# 2. Specifications

## 2.1 Flow Specification<sup>1</sup>

Parameter	Condition	SFM30	19	Units
Measurement range <sup>2</sup>		Air/O <sub>2</sub> : -	Air/O <sub>2</sub> : -10 to 240	
		Max.5	Typ <sup>4</sup>	
	span (0240slm)	±3%	±2%	m.v. <sup>6</sup>
Accuracy <sup>7</sup>	span (-100slm)	±5%	±3%	m.v. <sup>6</sup>
	Zero point	0.1 (max.)	0.05 (typ.)	slm <sup>3</sup>
A1 . 1 170	span	±1.5%	±1%	m.v. <sup>6</sup>
Noise Level <sup>7,8</sup>	Zero point	0.1		slm <sup>3</sup>
Span shift due to temperature variation		< 0.5% of rea	ading per 10°C	
Flow step response time (τ <sub>63</sub> )		< 3	ims	
Resolution		16	bit	
Calibrated for		O <sub>2</sub> , Air, mix	tures O <sub>2</sub> -air	
Media compatibility		Air, N <sub>2</sub> , O <sub>2</sub> , no	n-condensing	
Pressure Drop @60slm @200slm		<80	One mesh version <80 / 0.32 <500 / 2.0	

<sup>&</sup>lt;sup>1</sup> Unless otherwise noted, all sensor specifications are valid at 25°C with VDD = 3.3 V and absolute pressure = 966 mbar.

<sup>&</sup>lt;sup>2</sup> For other ranges contact Sensirion

<sup>&</sup>lt;sup>3</sup> In standard liter per minute at 20°C and 1013 mbar

<sup>&</sup>lt;sup>4</sup> For "Typ" a CpK of 0.67 is targeted (95% of sensors within the Typ limit)

<sup>&</sup>lt;sup>5</sup> for "Max" no sensor measured outside of these limits will be shipped and a CpK of 1.33 is targeted

<sup>&</sup>lt;sup>6</sup> Measured value

<sup>&</sup>lt;sup>7</sup> Span or offset value, whichever is greater

<sup>&</sup>lt;sup>8</sup> Noise Level is defined as standard deviation over individual sensor readings, measured in "average till read" mode with readout every 3ms. "Typ.": average of noise level, "Max." > 99% of sensors have a noise level below indicated value



# 2.2 Temperature Specification<sup>1</sup>

Parameter	Value
Measurement range	-20 °C to +85 °C
Resolution	16 bit
Accuracy	±2°C for range -10 °C to +60 °C, ±3°C otherwise
Repeatability	0.1°C

# 2.3 Electrical Specifications

Parameter	Symbol	Condition	Min.	Тур.	Max	Units	Comments
Supply Voltage	V <sub>DD</sub>		2.7	3.3	5.5	V	Recommended: 3.3V +/- 5% (calibration voltage)
Power-up/down level	VPOR		2.3	2.5	2.7	V	
		Measuring		3.8	5.5	mA	VDD 3.3V
Supply current	IDD	Idle state			1.1	mA	
		Sleep mode			1	uA	

## 2.4 Timing Specifications

Parameter	Symbol	Min.	Тур.	Max.	Units	Comments
Power-up time	t <sub>PU</sub>		2		ms	Time to sensor ready
Soft reset time	t <sub>sr</sub>			2	ms	Time between soft reset command or exit sleep mode and sensor ready
Warm-up time	t <sub>w</sub>		30		ms	To reach accuracy spec after first measurement command
I <sup>2</sup> C SCL frequency	f <sub>I2C</sub>		400	1000	kHz	
Update rate flow value		1800	2000	2200	Hz	
Update rate temperature value		112.5			Hz	Temperature value is updated at least every 16 flow values

<sup>&</sup>lt;sup>1</sup> The measured temperature is the temperature of the bulk silicon in the sensor. This temperature value is not only depending on the gas temperature, but also on the sensor's surroundings. Using the signal to measure solely the gas temperature will need special precautions, such as isolating the sensor from external temperate influences.



## 2.5 Mechanical Specifications

Parameter	Symbol	Min.	Тур.	Max.	Units	Condition/Comment
Operating pressure range		0.7		1.3	bar	absolute
Allowable overpressure	P <sub>max</sub>	-0.2		0.2	bar	gauge
Rated burst pressure	P <sub>burst</sub>			>1	bar	gauge
Weight	m			14.6	g	

### 2.6 Materials

Parameter	
	PPE+PS blend, Si, glass (Si <sub>3</sub> N <sub>4</sub> , SiO <sub>x</sub> ), gold, FR4, copper alloy, lead-free solder, epoxy, polyurethane, stainless steel (annealed)
REACH, RoHS	REACH and RoHS compliant

## 2.7 Absolute Minimum and Maximum Ratings

Parameter	Rating	Units
Supply Voltage V <sub>DD</sub>	-0.3 to 5.5	V
Max Voltage on pins (SDA, SCL)	-0.3 to V <sub>DD</sub> +0.3	V
Input current on any pin	±70	mA
Operating temperature range <sup>1</sup>	0 to +60	°C
Storage temperature range	-20 to +70	°C
Max. humidity for long term exposure	40	°C dew point
ESD HBM (human body model)	2	kV

## 2.8 Pin Assignment

The pin assignments of the SFM3019 series can be found in Table 1. The cap of the SFM3019 is compatible with DuraClik™ Wire-to-Board Receptacle Housing, Single Row, 4 Circuits. (Molex product number: 502351-0400).

Name	Description	digital:
SCL	Serial Clock (I <sup>2</sup> C Interface)	0.5
VDD	VDD Supply	Notations
GND	Connect to ground	
SDA	Bidirectional Serial Data (I <sup>2</sup> C Interface)	SDA GND VDD SCL

Table 1: SFM3019 series pin assignment.

<sup>&</sup>lt;sup>1</sup> For Air and N<sub>2</sub>. Long term exposure to (high concentrations of) O<sub>2</sub> at high temperatures can reduce the product lifetime.



#### 3. Measurement Mode

After the sensor receives a "start continuous measurement" command (for more details see Section 4.3.1) it enters the corresponding measurement mode and continuously performs a measurement every ~0.5ms. Therefore, a new reading can be obtained every 0.5ms. A single reading consists of three measurement values: flow, temperature, and one value for a status word (for more details on status word see Section 4.3.2).

### 3.1 Averaging of Flow Value

Two modes for on-sensor averaging of the flow signal are available:

- 1) Average-until-read: in this mode, the sensor averages all measured values prior to read out. This averaging mode is the default mode selected upon startup and after a reset.
- 2) Fixed-N averaging: in this mode, every reading is the average of a fixed number (N) of measured flow values.

### 3.1.1 Average-Until-Read

If the 'average till read' option is chosen, the sensor averages all values  $x_i$  prior to the read out. This has the benefit that the user can read out the sensor at his own desired speed, without losing information and thus preventing aliasing. During the first 64ms of averaging, the averaged value is obtained as the arithmetic mean:

$$\bar{x} = \sum_{i=1}^{M} \frac{x_i}{M} \quad \text{for } t < 64 \text{ ms}$$

If the period between readouts is longer than 64ms, the sensor will continue to average, but using a different algorithm. In this algorithm, exponential smoothing is used with a smoothing factor  $\alpha = 0.02$ :

$$S_k = \alpha \cdot x_k + (1 - \alpha) \cdot S_{k-1}$$
,  $S_0 = \bar{x}$ , for  $t > 64$  ms Where  $S_0$  is the

arithmetic mean value after the first 64ms and the readout value for flow is  $S_k$ .

With an exponential smoothing factor of  $\alpha$  = 0.02, the value read out by the user is an average value of about the last 64ms. In order not to lose information, it is recommended to read out the sensor at least once every 64ms. Please refer to the relevant literature for more information about exponential smoothing.

Average-until-read is the default averaging mode after startup and after a reset. It can further be selected by the user by setting *N*=0 in the argument for the "Configuration of Averaging" command (for more details see Section 4.3.5).

### 3.1.2 Fixed-N Averaging

Averaging may also be set to a fixed number  $1 \le N \le 128$  of measurements to be averaged (c.f. Section 4.3.5). This type of averaging is especially suited to avoid any averaging (N=1). If fixed-N averaging is chosen, the update time for new readings is N \* 0.5ms accordingly. Averaging has the benefit that the user can read out the sensor at his own desired speed, without losing information and thus preventing aliasing. In this case, the averaged value  $\bar{x}$  is the arithmetic mean of the individual,  $\sim 2kHz$  measurements x:

$$\bar{x} = \sum_{i=1}^{N} \frac{x_i}{N}$$

If no averaging is desired, set *N* to 1.



## 3.2 Sensor Start-Up and Warm-Up Behavior

The typical time for system power-up (until the sensor responds to communication requests) is 2ms. The typical time from a soft reset until the sensor responds to communication requests is also 2ms if the SCL line is high.

After reset or start-up of the sensor, the sensor's internal heater is off and is automatically turned on by performing a *Start Continuous Measurement* command (see Section 4.3.1). The very first measurement after *Start Continuous Measurement* is ready after approximately 12ms.

Due to the thermal measurement principle, a total warm-up time of typically 30ms is necessary for an accurate measurement. This includes the 12ms needed for measurement initialization.




Figure 1: Time diagram upon start-up.

### 3.3 Sensor Check at Start-Up

The gas flow sensor checks the integrity of its entire memory content (including all lookup tables for all gases and gas mixtures and the entire configuration) at start up automatically using a CRC check sum. In case the CRC check fails, the I<sup>2</sup>C-interface is deactivated.

The following command provides a further possibility for an integrity check: Read Product Identifier (command code 0xE102, see Section 4.3.9): Ideally suited to test if the sensor is connected correctly and if the sensor has been assigned the correct I<sup>2</sup>C-address upon start-up.



## 4. Digital Interface Description

The sensor's digital interface is compatible with the I<sup>2</sup>C protocol. This chapter describes the available command set. For detailed information about the I<sup>2</sup>C protocol, please consult the document "NXP I<sup>2</sup>C-bus specification and user manual" (http://www.nxp.com/documents/user\_manual/UM10204.pdf).

The physical interface consists of two bus lines: a data line (SDA) and a clock line (SCL) which need to be connected via pull-up resistors to the bus voltage of the system.


#### 4.1 I2C Address

The I<sup>2</sup>C address for SFM3019 in Hex and binary is 0x2E and b 0101110.

In the I<sup>2</sup>C protocol, a read or write bit follows the I<sup>2</sup>C address.


## 4.2 I<sup>2</sup>C Sequences

The commands have a length of 16 bits. If commands require an argument, they are followed by a 16-bit argument plus an 8bit checksum:



**Figure 2:** I<sup>2</sup>C sequences to send a command with an argument to the sensor.

After that, data is read from the sensor in multiples of 16-bit words, each followed by an 8-bit checksum to ensure communication reliability. I<sup>2</sup>C sequences can be aborted with a NACK and STOP condition.



**Figure 3:** I<sup>2</sup>C sequences to read results from the sensor. Dark areas with white text indicate that the sensor controls the SDA (Data) line.



### 4.3 I<sup>2</sup>C Commands

The command set consists of a set of various commands, which are described in the following sections:

- Various Start continuous measurement commands (one for each gas/gas mixture)
- Update concentration
- Stop continuous measurement command
- Configuration of averaging
- · Read Scale Factor, Offset, and Flow Unit
- Soft reset
- Entering and exiting sleep mode
- Read product identifier

#### 4.3.1 Start Continuous Measurement and Read Results

The sensor measures both flow and temperature and in a third data buffer a status word. All three measurement results can be read out through one single I<sup>2</sup>C read header when the continuous measurement is running. The specific command code used for the start continuous measurement command selects the calibrated gas or binary gas mixture (lookup table) for the flow signal.

Command code (Hex)	SFM3019	Command argument	Averaging
0x3603	Gas 0: O <sub>2</sub>	N/A	
0x3608	Gas 1: Air	N/A	
0x3615	Reserved for future use	N/A	
0x361E	Reserved for future use	N/A	According to Section.4.3.5
0x3624	Reserved for future use	N/A	
0x362F	Reserved for future use	N/A	
0x3632	Gas mixture 0: Air-O <sub>2</sub>	Volume fraction of O <sub>2</sub> (in ‰)	

Table 2: I<sup>2</sup>C commands to start continuous measurement

After the command has been sent, the chip continuously measures and updates the measurement results. New results (flow, temperature, and status word) can be read continuously with a single I<sup>2</sup>C read header for each measurement.

Further commands must not be sent until the stop measurement command has been sent. Consult section 4.3.4 for more details.

After the start measurement command is sent:

- the first measurement result will be available after 12ms
- small accuracy deviations (few % of reading) can occur during the first 30ms (including the 12ms)

If an I<sup>2</sup>C read header (I<sup>2</sup>C address + read bit) is sent to the sensor when no measurement data is available, the sensor will respond with a NACK condition.



Preceding command	Consecutive read	Description
Continuous measurement	Byte1: Flow 8msb Byte2: Flow 8lsb Byte3: CRC Byte4: Temp 8msb Byte5: Temp 8lsb Byte6: CRC Byte7: Status Word 8msb Byte8: Status Word 8lsb	After a start continuous measurement command, the measurement results can be read out.  The temperature and the consecutive bytes don't need to be read out (every time). The read sequence can be aborted by a NACK and a STOP condition.
	Byte9: CRC	

**Table 3:** Consecutive reads after I<sup>2</sup>C command to start continuous measurement. The flow, and temperature values are raw data value consisting of a 16bit signed integer. For conversion into physical units refer to sections 4.5.

#### 4.3.2 Status Word

The status word is returned along with every flow and temperature measurement. It contains the measurement command that is currently running, information on the currently selected averaging mode, and the defined gas concentration of the current measurement command (either a fixed value of 0x3FF for a pure calibration gas, or the volume fraction of  $O_2$  in per mille for gas mixtures). The table below lists the values returned by the status word.

Command Code of running measurement command	Bits <15:12>: currently running measurement command	Bit <11>: status of exponential smoothing	Bit <10>: averaging mode	Bits <9:0>: concentration
0x3603	0b0000			
0x3608	0b0001			Fixed value of 0b11111111111 = 0x3FF = 1023 to indicate that a pure gas is being measured
0x3615	0b0010	0: Exponential smoothing is not active  1: Averaging mode has switched to exponential smoothing	0: Average-Until-Read is active 1: Fixed-N averaging is active	
0x361E	0b0011			
0x3624	0b0100			
0x362F	0b0101			
0x3632	0b0110			
0x3639	0b0111			Volume fraction of Gas 0 (in ‰)
0x3646	0b1000			

**Table 4:** Values returned by the status word depending on the measurement command running. Rows, which are grayed out are reserved for future uses.

### 4.3.3 Update Concentration

To update the concentration of a binary gas mixture dynamically, i.e. without interrupting the running measurement mode, an I<sup>2</sup>C transmission sequence consisting of a set of two instructions is provided:

Command	Command code (Hex)	Command argument	Description
Transmit concentration	0xE17D	Volume fraction of Gas 0 in ‰	This instruction transmits the concentration value to the flow sensor.
Reset I <sup>2</sup> C address pointer	0xE000		This instruction resets the I <sup>2</sup> C address pointer to the regular result output buffer such that the measurement data as described in Sections 4.2 and 4.3 are obtained upon a subsequent read header.

**Table 5:** I<sup>2</sup>C commands to update the concentration for binary gas mixtures.



After those two commands changing the gas concentration, measurement results can be read with a read header as described in Section 4.3.1. No new start command is required.

#### Please note:

- The concentration value must not be updated more than once per millisecond.
- If a concentration value greater than 1000 is passed to the sensor, the measurement mode will stop.
- For the initial concentration value, which is supplied with the start measurement command (0x3632, c.f. Sec. 4.3.1) the command argument will be immediately applied without a need to send the second command.

Attention: Changing the gas mixture by using a new measurement command takes 12 ms to change the concentration. Therefore, the procedure to change the concentration only is significantly faster and should be used for dynamic adjustments of the concentration.

Important: During the transmission and execution of these two commands (in Table 5) no I<sup>2</sup>C-read must take place.

### 4.3.4 Stop Continuous Measurement

Command	Command code (Hex)	Description
Stop continuous measurement	0x3FF9	This command stops the continuous measurement and puts the sensor in idle mode. After it receives the stop command, the sensor needs up to 0.5ms to power down the heater, enter idle mode and be receptive for a new command.

**Table 6:** I<sup>2</sup>C command to stop continuous measurement

When the sensor is in continuous measurement mode, the sensor must be stopped before it can accept another command.

The only exceptions are the soft reset (see Section 4.3.7) command and the update concentration command (see Section 4.3.3)

In idle mode the sensor will consume less power, but consider the sleep mode for most effective energy saving results.

### 4.3.5 Configuration of Averaging

Command	Command code (Hex)	Command argument	Description
Configure averaging	0x366A	N	<ul> <li>This command configures the sensor's averaging mode:</li> <li>1) N=0 (default): average-until-read mode (c.f. Section 3.1)</li> <li>2) 1≤N≤128: fixed-N averaging mode. N is the number of internal measurements that are averaged for one returned measurement value (i.e. the average over N flow samples, where N = CmdArgument, c.f. Section 3.1).</li> <li>The configured averaging mode will be used for flow measurements until a reset or re-execution of this command is performed. After a reset, averaging is set to average-until-read mode (i.e. N is set to 0). The highest averaging number allowed is 128. If a higher number is used in the command argument, it will be overruled by the maximal value of 128 samples to average.</li> <li>If no averaging is desired, set N to 1.</li> </ul>

Table 7: I<sup>2</sup>C command to configure averaging



### 4.3.6 Read Scale Factor, Offset, and Flow Unit

This command provides the scale factor and offset to convert flow readings into physical units (see Section 4.5). The scale factor and offset are specific to the calibrated gas / gas mixture and its corresponding lookup table used for the flow measurement. Therefore, the gas / gas mixture needs to be specified in the command argument by the command code of the corresponding start continuous measurement (see Section 4.3.1).

For the SFM3019, the flow unit is a fixed value 0x0148 and corresponds to slm: standard liter per minute at 20°C and 1013mbar pressure.

Command	Command code (Hex)	Command argument (Hex)
Read Scale Factor, Offset, and Flow Unit	0x3661	Command code of desired gas / gas mixture for scale factor, offset and flow unit (Section 4.3.1). The command argument needs to be followed by CRC (Section 4.2)

Preceding command	Consecutive read	Description
Read Scale Factor, Offset, and Flow Unit	Byte2: Scale factor 8lsb Byte3: CRC	After a "Read Scale Factor, Offset, and Flow Unit" command, the corresponding results can be read out. The scale factor and offset are a 16-bit signed integer number represented by a two's complement (ranging from -32'768 to 32'767). The flow unit is a 16-bit identifier.

Table 8: I<sup>2</sup>C command to read the scale factor, offset, and flow unit. Result of this command.

### 4.3.7 Soft Reset

Command	I <sup>2</sup> C address + W bit + command code (Hex)		
General call reset	0x0006	NA	This sequence resets the sensor with a separate reset block, which is as much as possible detached from the rest of the system on chip.  Note that the I <sup>2</sup> C address is 0x00, which is the general call address, and that the command is 8 bits long. The reset is implemented according to the I <sup>2</sup> C specification.

Table 9: Reset command

After the reset command the sensor will typically take 2ms to reset. During this time the sensor will not acknowledge its address nor accept commands.

### 4.3.8 Entering and Exiting Sleep Mode

In sleep mode the sensor uses a minimum amount of power. The mode can only be entered from idle mode, i.e. when the sensor is not performing measurements.

This mode is particularly useful for battery operated devices. To minimize the current in this mode, the complexity of the sleep mode circuit has been reduced as much as possible, which is mainly reflected by the way the sensor exits the sleep mode. In sleep mode the sensor cannot be soft reset.



Command	Command code (Hex)	Consecutive read	Description
Enter Sleep mode	0x3677		The sleep command can be sent after a stop continuous measurement command has been issued and the sensor is in idle mode.
Exit Sleep mode	NA		The sensor exits the sleep mode and enters the idle mode when it receives the <b>valid I<sup>2</sup>C address and a write bit ('0').</b> Note that the I <sup>2</sup> C address is <b>not</b> acknowledged. It is necessary to poll the sensor to see whether the sensor has received the address and has woken up. This should take typically 16ms.

Table 10: Sleep mode commands

#### 4.3.9 Read Product Identifier

The product identifier and the serial number can be read out by sending the command below. The command can only be entered from the idle mode, i.e. when the sensor is not performing measurements.

Command	Command code (Hex)	Consecutive read	Description
Read product identifier	0xE102	Byte1: Product number [31:24] Byte2: Product number [23:16] Byte3: CRC Byte4: Product number [15:8] Byte5: Product number [7:0] Byte6: CRC Byte7: Serial number [63:56] Byte8: Serial number [55:48] Byte9: CRC Byte10: Serial number [47:40] Byte11: Serial number [39:32] Byte12: CRC Byte13: Serial number [31:24] Byte14: Serial number [23:16] Byte15: CRC Byte16: Serial number [15:8] Byte17: Serial number [7:0] Byte18: CRC	Note that the command need to be preceded with an I²C write header (I²C address + W).  The command returns:  - 32 bit unique product and revision number. The number is listed in Table 13 below.  Note that the last 8 bits are the revision number and are subject to change as long as the datasheet is preliminary.  - 64 bit unique serial number in the format of an unsigned long integer.  The serial number can be converted from binary into decimal, whereby in decimal it has the following format::  yywwxxxxxx, where: yy: last 2 digits of calibration year, ww: calibration week, xxxxxx: unique 6-digit sequential number within the calibration week.

Table 11: Read product identifier

Product	Product number	Comment
SFM3019	0x04020611	Last two digits:  8x indicating a prototype  1x indicating finalized products, therefore last two bytes expected to change during development

Table 12: Product number for SFM3019

## 4.4 Checksum Calculation

The 8-bit CRC checksum transmitted after each data word is generated by a CRC algorithm. Its properties are listed in Table 13. The CRC covers the contents of the two previously transmitted data bytes. To calculate the checksum, only these two previously transmitted data bytes are used.



Property	Value	
Name	CRC-8	
Protected data	read data	
Width	8 bit	
Polynomial	0x31 (x8 + x5 + x4 + 1)	
Initialization	0xFF	
Reflect input	False	
Reflect output	False	
Final XOR	0x00	
Example	CRC (0xBEEF) = 0x92	

Table 13: Checksum definition

## 4.5 Number Format and Conversion to Physical Values

The number format of the flow and temperature signals and the conversion to a physical value with a scale factor and an offset where applicable is explained below:

### 4.5.1 Scale Factors and Offsets

Cimal	SFM3019		0
Signal	Scale Factor	Offset	Comment
Gas 0: O <sub>2</sub>	170 slm <sup>-1</sup>	-24576	
Gas 1: Air	170 slm <sup>-1</sup>	-24576	Can be read out using an
Gas 2-5: reserved for future use	tbd	tbd	I <sup>2</sup> C-command (c.f. Sec.
Gas mixture 0: Gas1 - Gas0	170 slm <sup>-1</sup>	-24576	4.3.6)
Gas mixture 1: reserved for future use	tbd	tbd	,
Temperature	200 °C <sup>-1</sup>	0	

Table 14: Scale factors and offsets

#### 4.5.2 **Flow**

The digital calibrated gas flow signal read from the sensor is a 16-bit signed integer number represented by a two's complement (ranging from -32'768 to 32'767). The integer value can be converted to the physical value by subtracting the offset and dividing it by the scale factor (gas flow in  $slm = \frac{sensor\ output-offset}{scale\ factor}$ ). The scale factor and offset are specific to every calibrated gas / gas mixture / lookup table.

The flow unit slm signifies standard liters per minute with reference temperature equal to 20°C and reference pressure equal to 1013mbar.

### 4.5.3 **Temperature**

The digital calibrated temperature signal read from the sensor is a 16-bit signed integer number represented by a two's complement (ranging from -32'768 to 32'767). The integer value can be converted to the physical value by subtracting the offset and dividing it by the scale factor (temperature in  ${}^{\circ}C = \frac{Sensor\ Output - offset}{Scale\ factor}$ ).

#### 4.5.4 Flow Unit

The flow unit is specific to every calibrated gas / gas mixture / lookup table and is specified in Section 4.5.1. It can further be read out using an I<sup>2</sup>C-command (Section 4.3.6). The flow unit is given by a 16-bit unsigned word, where the information about the unit is encoded in the first 13 bits as follows:




- 1. Bits <3:0>: unit prefix (multiplier)
- 2. Bits <7:4>: time base (e.g. per minute)
- 3. Bits <12:8>: unit (e.g. standard liter)

The allowed values for the 3 unit constituents are given in the tables below:

Bits <12:8>	Signification: Unit	Bits <7:4>	Signification: Time Base	Bits <3:0>	Signification: Unit Prefix	overall code (16 bits)	Signification: flow unit
0b00001	standard liter	0b0100	per minute	0b1000	10-0 = 1	0b 0000 0001 0100	standard liter per
= 1	(T <sub>0</sub> =20°C,	= 4		= 8		1000	minute (slm) with
	p <sub>0</sub> =1013mbar)	·				= 0x0148	reference temperature =20°C and reference
						= 328	pressure =1013mbar

Table 15: Possible result(s) for the flow unit

# 5. Package Outline



# 6. Soldering

Standard wave soldering systems may be used for soldering SFMxxxx sensors. Reflow soldering is not feasible and may damage the sensor. Contact Sensirion for further information.



## 7. Label information

The SFM3019 Label on each sensor is composed of the following elements:

- Product part name (extended)
- Product order number
- Serial number (also as QR code)

QR code containing S/N SFM3019 – 240 3.000.400 S/N: XXXXXXXXXX

Figure 4: Label of the SFM3019

# 8. Shipping Package

- Packing units: 20 items/tray. Tray dimensions: 38.3 x 28.3 x 4.7 cm.
- Minimal order quantity: 6 trays. Box dimensions for 6 trays: 41 x 31 x 32 cm.

# 9. Revision History

Date	Author	Version	Changes
06.2021	PSIM		Removed preliminary and disclaimer. Completed footnotes. Flow specifications corrected.
07.2021	PSIM	v1.01	footnote correction



## 10. Important Notices

Warning, personal injury

Do not use this product as safety or emergency stop devices or in any other application where failure of the product could result in personal injury (including death). Do not use this product for applications other than its intended and authorized use. Before installing, handling, using or servicing this product, please consult the datasheet and application notes. Failure to comply with these instructions could result in death or serious injury.

If the Buyer shall purchase or use SENSIRION products for any unintended or unauthorized application, Buyer shall defend, indemnify and hold harmless SENSIRION and its officers, employees, subsidiaries, affiliates and distributors against all claims, costs, damages and expenses, and reasonable attorney fees arising out of, directly or indirectly, any claim of personal injury or death associated with such unintended or unauthorized use, even if SENSIRION shall be allegedly negligent with respect to the design or the manufacture of the product.

#### **ESD Precautions**

The inherent design of this component causes it to be sensitive to electrostatic discharge (ESD). To prevent ESD-induced damage and/or degradation, take customary and statutory ESD precautions when handling this product.

See application note "Handling Instructions" for more information.

#### Warranty

SENSIRION warrants solely to the original purchaser of this product for a period of 12 months (one year) from the date of delivery that this product shall be of the quality, material and workmanship defined in SENSIRION's published specifications of the product. Within such period, if proven to be defective, SENSIRION shall repair and/or replace this product, in SENSIRION's discretion, free of charge to the Buyer, provided that:notice in writing describing the defects shall be given to SENSIRION within fourteen (14) days after their appearance;

- such defects shall be found, to SENSIRION's reasonable satisfaction, to have arisen from SENSIRION's faulty design, material, or workmanship; 

  the defective product shall be returned to SENSIRION's factory at the Buyer's expense; and
- the warranty period for any repaired or replaced product shall be limited to the unexpired portion of the original period.

This warranty does not apply to any equipment which has not been installed and used within the specifications recommended by SENSIRION for the intended and proper use of the equipment. EXCEPT FOR THE WARRANTIES EXPRESSLY SET FORTH HEREIN, SENSIRION MAKES NO WARRANTIES, EITHER EXPRESS OR IMPLIED, WITH RESPECT TO THE PRODUCT. ANY AND ALL WARRANTIES, INCLUDING WITHOUT LIMITATION, WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE, ARE EXPRESSLY EXCLUDED AND DECLINED.

SENSIRION is only liable for defects of this product arising under the conditions of operation provided for in the datasheet and proper use of the goods. SENSIRION explicitly disclaims all warranties, express or implied, for any period during which the goods are operated or stored not in accordance with the technical specifications.

SENSIRION does not assume any liability arising out of any application or use of any product or circuit and specifically disclaims any and all liability, including without limitation consequential or incidental damages. All operating parameters, including without limitation recommended parameters, must be validated for each customer's applications by customer's technical experts. Recommended parameters can and do vary in different applications.

SENSIRION reserves the right, without further notice, (i) to change the product specifications and/or the information in this document and (ii) to improve reliability, functions and design of this product.

Copyright © 2020, SENSIRION. CMOSens® is a trademark of Sensirion All rights reserved

## **Headquarters and Subsidiaries**

Sensirion AG

Laubisruetistr. 50 CH-8712 Staefa ZH Switzerland

phone: +41 44 306 40 00 fax: +41 44 306 40 30 info@sensirion.com
www.sensirion.com

Sensirion Taiwan Co. Ltd phone: +886 3 5506701 info@sensirion.com www.sensirion.com Sensirion Inc., USA

phone: +1 312 690 5858 info-us@sensirion.com www.sensirion.com

Sensirion Japan Co. Ltd. phone: +81 3 3444 4940 info-jp@sensirion.com www.sensirion.co.jp Sensirion Korea Co. Ltd. phone: +82 31 337 7700~3 info-kr@sensirion.com www.sensirion.co.kr

Sensirion China Co. Ltd. phone: +86 755 8252 1501 info-cn@sensirion.com www.sensirion.com.cn

To find your local representative, please visit <a href="https://www.sensirion.com/distributors">www.sensirion.com/distributors</a>

# **X-ON Electronics**

Largest Supplier of Electrical and Electronic Components

Click to view similar products for Environmental Test Equipment category:

Click to view products by Sensirion manufacturer:

Other Similar products are found below:

CW40 F150C10E3DRT F150CD10E2 F150L75 F150LRS 2837806 4328074 4366444 S-11 382153 700PTP-1 GEO-CABLE-REEL-50M H115 H300 F150-SLC50 AW-CO-1000 AW-H2S-100 AW-NmHc-100 ES1-CO-1000-01 ES4-CO-1000-01 TM-414 HS115 SHG-02 SHS-05 SHG-01 RH520-220 TESTO 175-H1 0572 1754 TESTO 830-T1 0560 8311 AX-5002 AX-5003 AX-B150 AX-B180 AX-B350 AX-L230 AX-PH02 12227021 12228657 12229067 12230090 7358958 12226653 12230146 12230171 12229190 12228801 12228665 12228819 12229360 12230502 12229378